

Dr Sanjana Shastri¹, Kiran Rote², Anita Sawant², Dr Rushika Shah¹, Dr Meena Lopez¹, Dr Senthilnathan M¹, Dr Jaideep Gogtay¹

		Formoterol			Glycopyrronium			Budesonide		
		Assay % L.C	UDD % TDD	% FPF	Assay %L.C	UDD % TDD	% FPF	Assay %L.C	UDD % TDD	% FPF
Brand and Device	Batch No.									

Cipla FGB Synchronobreathe	BA12360	100	98	53	99	99	47	97	95	53
Cipla FGB pMDI	BA12360	101	96	54	106	94	50	103	93	46
Innovator (Breztri Aerosphere)	6100878C00	113	104	49	102	97	50	103	96	46

Conclusion: The in-vitro performance of Cipla's MDI and Synchronobreathe is within the accepted limits and comparable to the innovator i.e. Breztri Aerosphere.